#### §878.4040

as growth factors, or is composed of materials derived from animal sources.

(b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in part 807, subpart E of this chapter subject to the limitations in §878.9.

[64 FR 53929, Oct. 5, 1999]

#### §878.4040 Surgical apparel.

- (a) Identification. Surgical apparel are devices that are intended to be worn by operating room personnel during surgical procedures to protect both the surgical patient and the operating room personnel from transfer of microorganisms, body fluids, and particulate material. Examples include surgical caps, hoods, masks, gowns, operating room shoes and shoe covers, and isolation masks and gowns. Surgical suits and dresses, commonly known as scrub suits, are excluded.
- (b) Classification. (1) Class II (special controls) for surgical gowns and surgical masks.
- (2) Class I (general controls) for surgical apparel other than surgical gowns and surgical masks. The class I device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to § 878.9.

[53 FR 23872, June 24, 1988, as amended at 65 FR 2317, Jan. 14, 2000]

### §878.4100 Organ bag.

- (a) *Identification*. An organ bag is a device that is a flexible plastic bag intended to be used as a temporary receptacle for an organ during surgical procedures to prevent moisture loss.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to §878.9.
- [53 FR 23872, June 24, 1988, as amended at 59 FR 63010, Dec. 7, 1994; 65 FR 2318, Jan. 14, 2000]

## §878.4160 Surgical camera and accessories.

- (a) *Identification*. A surgical camera and accessories is a device intended to be used to record operative procedures.
- (b) Classification. Class I. The device is exempt from the premarket notifica-

tion procedures in subpart E of part 807 of this chapter.

 $[53~{\rm FR}~23872,~{\rm June}~24,~1988,~{\rm as~amended}~{\rm at}~54~{\rm FR}~13827,~{\rm Apr.}~5,~1989]$ 

# §878.4200 Introduction/drainage catheter and accessories.

- (a) Identification. An introduction/ drainage catheter is a device that is a flexible single or multilumen tube intended to be used to introduce nondrug fluids into body cavities other than blood vessels, drain fluids from body cavities, or evaluate certain physiologic conditions. Examples include irrigation and drainage catheters, pediatric catheters, peritoneal catheters (including dialysis), and other general surgical catheters. An introduction/ drainage catheter accessory is intended to aid in the manipulation of or insertion of the device into the body. Examples of accessories include adaptors, connectors, and catheter needles.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to §878.9.

[53 FR 23872, June 24, 1988, as amended at 65 FR 2318, Jan. 14, 2000]

### §878.4300 Implantable clip.

- (a) *Identification*. An implantable clip is a clip-like device intended to connect internal tissues to aid healing. It is not absorbable.
  - (b) Classification. Class II.

#### §878.4320 Removable skin clip.

- (a) *Identification*. A removable skin clip is a clip-like device intended to connect skin tissues temporarily to aid healing. It is not absorbable.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to §878.9.

[53 FR 23872, June 24, 1988, as amended at 65 FR 2318, Jan. 14, 2000]

## § 878.4350 Cryosurgical unit and accessories.

(a) Identification—(1) Cryosurgical unit with a liquid nitrogen cooled cryoprobe and accessories. A cryosurgical unit with a liquid nitrogen cooled cryoprobe

#### Food and Drug Administration, HHS

and accessories is a device intended to destroy tissue during surgical procedures by applying extreme cold.

- (2) Cryosurgical unit with a nitrous oxide cooled cryoprobe and accessories. A cryosurgical unit with a nitrous oxide cooled cryoprobe and accessories is a device intended to destroy tissue during surgical procedures, including urological applications, by applying extreme cold.
- (3) Cryosurgical unit with a carbon dioxide cooled cryoprobe or a carbon dioxide dry ice applicator and accessories. A cryosurgical unit with a carbon dioxide cooled cryoprobe or a carbon dioxide dry ice applicator and accessories is a device intended to destroy tissue during surgical procedures by applying extreme cold. The device is intended to treat disease conditions such as tumors, skin cancers, acne scars, or hemangiomas (benign tumors consisting of newly formed blood vessels) and various benign or malignant gynecological conditions affecting vulvar, vaginal, or cervical tissue. The device is not intended for urological applica-
  - (b) Classification. Class II.

# § 878.4370 Surgical drape and drape accessories.

- (a) Identification. A surgical drape and drape accessories is a device made of natural or synthetic materials intended to be used as a protective patient covering, such as to isolate a site of surgical incision from microbial and other contamination. The device includes a plastic wound protector that may adhere to the skin around a surgical incision or be placed in a wound to cover its exposed edges, and a latex drape with a self-retaining finger cot that is intended to allow repeated insertion of the surgeon's finger into the rectum during performance of a transurethral prostatectomy.
  - (b) Classification. Class II.

#### §878.4380 Drape adhesive.

- (a) *Identification*. A drape adhesive is a device intended to be placed on the skin to attach a surgical drape.
- (b) Classification. Class I. The device is exempt from the premarket notifica-

tion procedures in subpart E of part 807 of this chapter.

[53 FR 23872, June 24, 1988, as amended at 59 FR 63010, Dec. 7, 1994]

# §878.4400 Electrosurgical cutting and coagulation device and accessories.

- (a) *Identification*. An electrosurgical cutting and coagulation device and accessories is a device intended to remove tissue and control bleeding by use of high-frequency electrical current.
  - (b) Classification. Class II.

#### §878.4440 Eye pad.

- (a) *Identification*. An eye pad is a device that consists of a pad made of various materials, such as gauze and cotton, intended for use as a bandage over the eye for protection or absorption of secretions.
- (b) Classification. Class I. The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter.
- [53 FR 23872, June 24, 1988, as amended at 59 FR 63010, Dec. 7, 1994]

## § 878.4450 Nonabsorbable gauze for internal use.

- (a) Identification. Nonabsorbable gauze for internal use is a device made of an open mesh fabric intended to be used inside the body or a surgical incision or applied to internal organs or structures, to control bleeding, absorbluid, or protect organs or structures from abrasion, drying, or contamination. The device is woven from material made of not less than 50 percent by mass cotton, cellulose, or a simple chemical derivative of cellulose, and contains x-ray detectable elements.
- (b) Classification. Class I. The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter.

 $[53\ \mathrm{FR}\ 23872,\ \mathrm{June}\ 24,\ 1988,\ \mathrm{as}\ \mathrm{amended}\ \mathrm{at}\ 61\ \mathrm{FR}\ 1123,\ \mathrm{Jan}.\ 16,\ 1996]$ 

### §878.4460 Surgeon's glove.

(a) *Identification*. A surgeon's glove is a device made of natural or synthetic rubber intended to be worn by operating room personnel to protect a surgical wound from contamination. The